In-Office Biological Monitoring System - Record Keeper

Systéme de surveillance biologique - Journal d'enregistrement

Initials *Initiales*	**Sterilizer Number & Type** *Numéro et type de sterilizateur*	**Date & Time into Incubator** *Date et heure dans l'incubateur*	**Date & Time out of Incubator** *Date et heure á l'exterieur de l'incubateur*	*** Results (Circle One)** ** Résultats (En encercler un)*	**** Control (Circle One)** *** Contrôle (En encercler un)*
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)
				PASS (Purple) **FAIL** (Yellow)	**PASS** (Yellow) **FAIL** (Purple)

* A sterilized biological indicator should remain purple after incubation, indicating adequate sterilization.
** The control biological indicator should turn yellow, indicating spore growth and a viable test.

** Un bioindicateur stérilisé doit demeurer violet aprés l'incubation, ce qui indique qu'il y a eu une stérilisation adéquate.*
*** Le bioindicateur de contrôle doit devenir jaune, ce qui indique qu'il y a un développement des spores et un test viable.*

Customer Care
1-800-819-3336
customercare@crosstex.com
*Visit **cdc.gov** for additional information*

Infection Control Helpline
TOLL FREE (1-855-878-3745)
1-8558-STERILE

CROSSTEX A CANTEL MEDICAL COMPANY | PROTECTS®