

Table 1

Contents of a Sterilization Log Book

Date
Autoclave # (if more than 1 sterilizer) or Type
Batch # (which load of the day)
Operator Initials
Time/Temperature Details
CI Strip Details (Pass/Fail)
Spore Test Details (Pass/Fail)
Load Contents

Sterilization Log #1

Machine Name: Midmark M11 Serial # V1571982 Clinic Dental

Date	Batch #	Operator initials	Load Contents:	Temp time verified Y/N	Indicator Strip Result Pass/Fail	BI/Spore test performed Y/N	Failed BI/Spore or Indicator Strip Actions:
3/17/20	2	AK	Handpieces - 3 file - Basic - 1 Rest - Isolate - 1 Hyg - Bur block - 2 1-mouth 2 - hygiene cass prop	Y	P	N	
3/17/20	3	AK	1 bur block set 1 syringe 2 rest cass 4 HLL Handpiece 1 matrix 1 basic	Y	P	N	
3/17/20	4	KB	2 Hyg Cass # HLL hand piece 1 bur set 2 Separators 1 memo.	Y	P	N	
3/17/20	1	KB	1 Rest Cass 1 Propri angle 1 Hyg. Cass 1 ED Frame 2 CT hand piece 1 bur set 1 cav. Tr.P	Y	P	N	
3/17/20	2		1 endo Cass 1 Dr. L bur 1 endo Cass/Pin/Halter 2 Basic 1 High/slow 1 endo slow	Y	P	N	
3/17/20	1	PR	1 Rest Cass 1 syng 1 Basic 1 pin/HLL 1 DR L bur 1 Hyg. Cass	Y	P	N	

Biological Record Sheet

Test Date	Batch #	Operator	Test Results	Control Results	Comments
3/17/20	1	AK	Pass	Pass	
3/17/20	2	AK	Pass	Pass	
3/17/20	3	AK	Pass	Pass	
3/17/20	4	KB	Pass	Pass	
3/17/20	1	KB	Pass	Pass	
3/17/20	2		Pass	Pass	

Biological Results: - No color change in media (sterile) + Media changes to yellow (non-sterile)